

# Calibration certificate for high temperatures



The loggers that can be provided on request or are always provided with a calibration certificate for temperature are:

- SterilDisk series
- Micro series
- MicroW series (some versions with particular certificates)
- S-Disk J series (some versions with particular certificates)
- Radio series

Standard calibration points are:

- 25°C
- 50°C
- 75°C
- 100°C (last point for Pasteur- and P- series)
- 125°C
- 140°C

Temperatura di riferimento / Reference temperature	Temperatura calcolata / Calculated temperature	Temperatura del sensore / Sensor temperature
25.75	25.75	0.00
50.00	50.00	0.00
75.00	75.00	0.00
100.00	100.00	0.00
125.00	125.00	0.00
140.00	140.00	0.00

## Main features

