

Calibration certificate of humidity



The loggers that can be provided on request or are always provided with a calibration certificate for humidity are:

- HumiStick
- HumiUSB series
- HumiPDF
- RHTPUSB
- PressPDF
- RHTemp series

Standard calibration points, not for RHTemp series, are:

- 5°C
- 30°C
- 50°C
- 30% @ 25°C
- 50% @ 25°C
- 70% @ 25°C

Standard calibration points, for RHTemp series, are:

- 25°C
- 50°C
- 80°C
- 125°C (RHTemp 125 only)
- 30% @ 25°C
- 50% @ 25°C
- 70% @ 25°C

Certificato di verifica calibratore
Certificate of calibration verification

Strumento / Instrument: Data Logger di temperatura e umidità / Temperature and Humidity Data Logger
Costruttore / Manufacturer: Tecnosoft
Modello / Model: HumiStick
Numero di serie / Serial Number: 1000028000740

Temperatura di riferimento / Reference temperature °C	Temperatura misurata / Temperature measured °C	Scostamento rispetto al riferimento / Error
0.00	0.00	0.00
50.00	50.00	0.00
125.00	125.00	0.00

Umidità nominale / Nominale humidity %RH	Umidità registrata / Measured humidity %RH	Scostamento rispetto al riferimento / Error
30.00	30.00	0.00
50.00	50.00	0.00
70.00	70.00	0.00

Modello in data / Printed on: 01/08/2018
Data di calibratura / Calibration date: 01/08/2018 YAMA / YAMA: 01/08/2011
Camera climatica / Climate chamber: Climatic Chamber Model: MIP 240-20-10002

Main features



Tecnosoft srl
Via Galvani, 4, 20068, Peschiera Borromeo (MI), Italy
T: (+39) 02 2692 2888 - F: (+39) 02 2692 2875
email: info@tecnosoft.eu - web: www.tecnosoft.eu
UNI EN ISO 9001:2008 Certiquality/IQNet N. 17733

