

Calibration certificate of pressure



The loggers that are always provided with a calibration certificate for temperature and/or pressure are:

- PressureDisk

Standard calibration points are:

- 25°C
- 50°C
- 75°C
- 100°C
- 125°C
- 140°C
- 10 mbar
- 30 mbar (PressureDisk V only)
- 100 mbar (PressureDisk V only)
- 300 mbar (PressureDisk V only)
- 1 bar (not for PressureDisk V)
- 2 bar (not for PressureDisk V)
- 3 bar (not for PressureDisk V)
- 4 bar (not for PressureDisk V)

Certificato di verifica calibrations / Calibration certificate

Modello / Model: PressureDisk
Numero di serie / Serial number: 100 000001

Temperature / Temperature

Temperatura di riferimento / Reference temperature	Temperatura Calibrata / Calibrated temperature	Scostamento sul nominale / Error
25.0	25.0	0.00
50.0	50.0	0.00
75.0	75.0	0.00
100.0	100.0	0.00
125.0	125.0	0.00
140.0	140.0	0.00

Scostamento sul nominale di calibrations / Calibration system accuracy: 0.1 °C

Pressure / Pressure

Pressione di riferimento / Reference pressure	Pressione Calibrata / Calibrated pressure	Scostamento sul nominale / Error
10.0	10.0	0.00
30.0	30.0	0.00
100.0	100.0	0.00
300.0	300.0	0.00

Camera di calibrazione / Calibration chamber: Brno Climate Chamber Model MHP 200 00108 01000
Atmosfera di pressione / Pressure reference: 20°C / 1013.25 hPa / Calibration Certificate No. 001-00-10 issued by Accredia N. 714 on 10 December 2011

Main features

