

Calibration certificate for high temperatures



The loggers that can be provided on request or are always provided with a calibration certificate for temperature are:

- SterilDisk series
- Micro series
- MicroW series (some versions with particular certificates)
- S-Disk J series (some versions with particular certificates)
- Radio series

Standard calibration points are:

- 25°C
- 50°C
- 75°C
- 100°C (last point for Pasteur- and P- series)
- 125°C
- 140°C

TECNO Soft
S.p.A. - Via Galvani, 4 - 20068, Peschiera Borromeo (MI) - Italy
Tel. +39 02 2692 2888 - Fax +39 02 2692 2875
Email: info@tecnosoft.eu - Web: www.tecnosoft.eu

Certificato di calibrazione / Calibration certificate

Strumento / Instrument: Data logger di temperatura / Temperature data logger
Distributore / Manufacturer: TECNO Soft
Modello / Model: 100000000
Numero di serie / Serial number: 000000000
Numero certificato / Certificate number: 000000000

Temperatura di riferimento Reference temperature °C	Temperatura calcolata Calculated temperature °C	Scostamento dalla temperatura Offset °C
25.75	25.75	0.00
50.00	50.00	0.00
75.00	75.00	0.00
100.00	100.00	0.00
125.00	125.00	0.00
140.00	140.00	0.00

Data di calibrazione / Calibration date: 2009/09/10
Data interrogazione precedente calibrazione /
Interrogation date for next calibration: 2009/09/11
Strumento di calibrazione / Calibration instrument: Siku Thermobath Bath CT0600 000 111 00007
Termometro campione / Reference thermometer: Phipps-Robinson Resistance, standard 10

Technosoft Data logger 100000000 is an ISO9001:2008 certified calibration certificate of 100000000

