

Calibration certificate for standard temperature

The loggers that can be provided on request or are always provided with a calibration certificate for temperature are:

- TempStick series (some versions with particular certificates)
- intelligent sensors series (some versions with particular certificates)
- TempUSB series
- TempNFC series
- TempPDF series
- TempBLE series

Standard calibration points are:

- -30°C
- 0°C
- 20°C
- 60°C (but TempNFC whom highest calibration point is 50°C)

Certificato di calibrazione
Calibration certificate

Strumento / Instrument: Data Logger di temperatura / Temperature Data Logger

Costruttore / Manufacturer: Secordsoft srl

Modello / Model: Tempstick

Numero di serie / Serial number: 0000020002940

Temperatura di riferimento Reference temperature °C	Temperatura misurata Measured temperature °C	Temperatura calibrata Calibrated temperature °C	Scostamento dal riferimento Offset °C
-30.00	-29.99	-30.00	0.00
-5.00	-5.00	-5.00	0.00
0.00	0.00	0.00	0.00
20.00	20.00	20.00	0.00
60.00	60.00	60.00	0.00

Data della stampa / Date of print: 08/11/2013
Data di calibrazione / Calibration date: 08/11/2013
Scadenza calibrazione / Calibration expiry: 08/05/2015
Strumento di calibrazione / Calibration instrument: Breda MPF 200 00.00002
Termometro campione / Reference thermometer: Platinum Resistence Delta Open pin 10002006