


# Calibration certificate for standard temperature



The loggers that can be provided on request or are always provided with a calibration certificate for temperature are:

- TempStick series (some versions with particular certificates)
- intelligent sensors series (some versions with particular certificates)
- TempUSB series
- TempNFC series
- TempPDF series
- TempBLE series

Standard calibration points are:

- -30°C
- 0°C
- 20°C
- 60°C (but TempNFC whom highest calibration point is 50°C)

**CERTIFICATO DI CALIBRAZIONE**  
**CALIBRATION CERTIFICATE**

Strumento / Instrument: Data Logger di temperatura / Temperature Data Logger  
Costruttore / Manufacturer: Scandsoft srl  
Modello / Model: TempStick  
Numero di serie / Serial number: 0000020002040

Temperatura di riferimento Reference temperature	Temperatura misurata Measured temperature	Temperatura calcolata Calculated temperature	Residuo della taratura Residual
-30.00	-30.01	-30.00	0.01
0.00	0.00	0.00	0.00
20.00	20.01	20.00	0.01
60.00	60.01	60.00	0.01

Data della stampa / Date of print: 08/11/2013  
Data di calibrazione / Calibration date: 08/11/2013  
Stazione di calibrazione / Calibration setup: 0005/0015  
Strumento di calibrazione / Calibration instrument: Sonda RTD 200 20.0000  
Termometro campione / Reference thermometer: Platinum Resistence Delta Open 1000200

## Main features

