Calibration certificate of pressure

The loggers that are always provided with a calibration certificate for temperature and/or pressure are:

- PressureDisk

Standard calibration points are:

- 25°C
- 50°C
- 75°C
- 100°C
- 125°C
- 140°C
- 10 mbar
- 30 mbar (PressureDisk V only)
- 100 mbar (PressureDisk V only)
- 300 mbar (PressureDisk V only)
- 1 bar (not for PressureDisk V)
- 2 bar (not for PressureDisk V)
- 3 bar (not for PressureDisk V)
- 4 bar (not for PressureDisk V)



Certificato di verifica calibrazione / Calibration certificate

Tecnosoft srl
Via Galvani, 4, 20068, Peschiera Borromeo (MI), Italy
T: (+39) 02 2692 2888 - F: (+39) 02 2692 2875
email: info@tecnosoft.eu - web: www.tecnosoft.eu
UNI EN ISO 9001:2008 Certiquality/IQNet N. 17733