

Calibration certificate of humidity



The loggers that can be provided on request or are always provided with a calibration certificate for humidity are:

- HumiStick
- HumiUSB series
- HumiPDF
- RHTEMP
- PressPDF
- RHTemp series

Standard calibration points, not for RHTemp series, are:

- 5°C
- 30°C
- 50°C
- 30% @ 25°C
- 50% @ 25°C
- 70% @ 25°C

Standard calibration points, for RHTemp series, are:

- 25°C
- 50°C
- 80°C
- 125°C (RHTemp 125 only)
- 30% @ 25°C
- 50% @ 25°C
- 70% @ 25°C

Calibration Certificate
Certificato di verifica calibratura
Certificate of calibration verification

Strumento / Instrument: Data Logger di temperatura e umidità / Temperature and Humidity Data Logger
Costruttore / Manufacturer: Tecnosoft
Modello / Model: HumiStick
Numero di serie / Serial Number: 1000028000740

Temperatura di riferimento Reference temperature °C	Temperatura misurata Temperature measured °C	Scostamento rispetto al riferimento Error
5.00	4.99	0.01
30.00	30.00	0.00
50.00	49.99	0.01

Umidità relativa relative humidity standard %RH	Umidità relativa relative humidity measured %RH	Scostamento rispetto al riferimento Error
30.00	30.00	0.00
50.00	50.00	0.00
70.00	70.00	0.00

Umidità misurata a temperatura di 25°C
Humidity measured at 25°C

Stampato in data / Printed on: 01/08/2018
Data di calibratura / Calibration date: 01/08/2018
Camera climatica / Climate chamber: YAMAHA / YAMAHA 01082017
Camera climatica / Climate chamber: Climatic Chamber Model MPP 240-30-10002

