

Battery kit for RHTemp80 L and RHTemp125 L

RHTemp80 L, RHTemp125 L battery is easily and quickly user replaceable too. New battery kits can be purchased (each kit includes a new battery, a new o-ring that must always be replaced whenever logger is opened and substitution instructions) or the service for battery replacement at Tecnosoft, usually along with recalibration.

Note: if used below -30°C , battery life will be highly reduced.



Accessories

- Universal tool for opening logger

